



INTENDED USE:

For *in vitro* diagnostic use.

Tissue-Tek Genie® anti-CD63 Rabbit Monoclonal Antibody [EP211] is designed to qualitatively detect CD63 protein in Formalin-Fixed, Paraffin Embedded (FFPE) specimen sections by immunohistochemical (IHC) staining on the Tissue-Tek Genie® Advanced Staining System. The clinical interpretation must be made in conjunction with histological examination, relevant clinical information, other diagnostic tests, and review of proper controls by a qualified pathologist.

Limitations

This product has been optimized for use with the default protocol for this antibody on the Tissue-Tek Genie Advanced Staining System, using Tissue-Tek Genie reagents and FFPE specimen sections. Staining quality may diminish when used with other systems and/or reagents.

SUMMARY & PRINCIPLE:

CD63 is a 53 kD lysosomal membrane protein expressed by a broad range of tissues/cell types, including platelets, granulocytes, basophils, a subset of T-cells, endothelial cells, sweat gland, islets of Langerhans, pancreas, prostate, and melanoma. The antibody labels CD63 protein in normal and neoplastic cells. It is a useful tool for identifying/classifying neoplasia such as melanoma, Merkel cell carcinoma, renal cell carcinoma, and adenocarcinoma of the breast and lung, when used with a panel of antibodies.

The Tissue-Tek Genie anti-CD63 Rabbit Monoclonal Antibody [EP211] is a primary antibody against the human CD63 protein and is provided in buffered saline containing 1% bovine serum albumin and 0.09% sodium azide. FFPE specimen sections are placed on positively charged slides and the paraffin is removed using the Tissue-Tek Genie® Dewax Solution (REF 8865),

after which heat-induced epitope retrieval is performed using the Tissue-Tek Genie® High pH Antigen Retrieval Solution (REF 8744). IHC demonstration of CD63 protein in FFPE specimen sections is achieved through use of the Tissue-Tek Genie anti-CD63 Rabbit Monoclonal Antibody [EP211] and the Tissue-Tek Genie® Pro Detection Kit, DAB (REF 8826). This procedure entails the sequential application of antibody and kit components as follows:

- Tissue-Tek Genie® Protein Block
- Tissue-Tek Genie® anti-CD63 Rabbit Monoclonal Antibody [EP211]
- Tissue-Tek Genie® Peroxidase Block
- Tissue-Tek Genie® Link (binds to the primary antibody)
- Tissue-Tek Genie® Polymer HRP-Conjugate (binds to the Link)
- Tissue-Tek Genie® DAB Substrate (visualizes the detected protein)

Tissue-Tek Genie® Hematoxylin (REF 8830) is then used to visualize the nuclei of cells. The IHC stained slide is cover-slipped and the FFPE specimen section reviewed using a light microscope.

EXPECTED RESULTS:

Specificity and intended use of this antibody were validated by performing IHC staining on the Tissue-Tek Genie Advanced Staining System using FFPE normal and tumor specimen sections. Cytoplasmic staining is observed with acinar cells in pancreas, melanocytes in skin, epithelium in prostate, and some melanomas.

Sensitivity and identification of the CD63 protein by this antibody may be affected by improper tissue handling. This may alter antigenicity, weaken detection and may generate false negative results.

Cellular Staining Pattern: Cytoplasmic

Positive Tissue Control: Pancreas, Skin, Prostate, Melanoma



CAUTIONS & WARNINGS:

For professional use only. Take reasonable precautions when handling. Avoid contact of reagents with eyes, skin, and mucous membranes. Wear protective gloves, clothing, and eye/face protection.

Capsules filled with ready-to-use, pre-diluted, antibody are for single use only. Do not attempt to re-fill or add additional reagent. Discard Capsule after use.

Cartridges filled with ready-to-use, pre-diluted antibody are intended for multiple uses. Do not attempt to re-fill or add additional reagent. Discard Cartridge when empty.

It is recommended to include appropriate controls on each specimen slide to help in identifying any deviation that might occur during the staining process.

All disposal practices must be in compliance with all Federal, State/Provincial and local laws and regulations. Refer to the SDS for further information.

STORAGE CONDITIONS:

Store this product at 2-8°C.

INSTRUCTIONS FOR USE:

Tissue-Tek Genie® anti-CD63 Rabbit Monoclonal Antibody [EP211], Capsules (REF 8283-C010):

1. Place the Reagent Dispensing Area (RDA)-Tag attached to the Capsule into the RDA.
2. Push the Capsule into the RDA with foil side down.
3. Place the RDA on the desired station of the Tissue-Tek Genie Advanced Staining System.
4. Place the slide with the specimen section on the same station, specimen section side down.
5. Assign Protocol 8283 to the same station.
6. Initiate execution of Protocol 8283.

7. The RDA-Tag 8283 will be scanned and registered automatically when the staining process is initiated.

8. During the primary antibody application step, the antibody will be released from the Capsule into the RDA to the specimen section on the slide.

9. The staining Protocol continues to the end.

Tissue-Tek Genie® anti-CD63 Rabbit Monoclonal Antibody [EP211], Cartridge (REF 8283-M250):

1. Prior to placing the Cartridge on the Carousel of the Tissue-Tek Genie Advanced Staining System, prime the Cartridge by facing the nozzle downwards and gently pinching the nozzle tubing until the tubing is filled with the reagent.
2. Place the Cartridge on the Carousel.
3. Place the RDA-Tag 8283 into the RDA.
4. Place the RDA on the desired station of the Tissue-Tek Genie Advanced Staining System.
5. Place the slide with the specimen section on the same station, specimen section side down.
6. Assign Protocol 8283 to the same station.
7. Initiate execution of Protocol 8283.
8. The RDA-Tag 8283 and the Cartridge will be scanned and registered automatically when the staining process is initiated.
9. During the primary antibody application step, the antibody will be dispensed from the Cartridge into the RDA to the specimen section on the slide.
10. The staining Protocol continues to the end.

Materials Required but Not Supplied

The following reagents may be required for staining but are not provided:

- Tissue-Tek Genie® Dewax Solution (REF 8865)
- Tissue-Tek Genie® Wash Solution (REF 8874)
- Tissue-Tek Genie® High pH Antigen Retrieval Solution (REF 8744)
- Tissue-Tek Genie® Non-immune Rabbit Ig Antibody, Negative Control (REF 8605)

Tissue-Tek Genie® anti-CD63 Rabbit Monoclonal Antibody [EP211]



- Tissue-Tek Genie® *Pro* Detection Kit, DAB (REF 8826)
- Tissue-Tek Genie® Hematoxylin (REF 8830)

Further information can be found on the Sakura Finetek USA website at www.sakura-americas.com/Genie

ORDER INFORMATION:

Product Code/Product Name

REF 8283-C010 Tissue-Tek Genie® anti-CD63 Rabbit Monoclonal Antibody [EP211], 10 Capsules, Ready-To-Use.

REF 8283-M250 Tissue-Tek Genie® anti-CD63 Rabbit Monoclonal Antibody [EP211], 1 Cartridge, 250 Tests, Ready-To-Use.

NOTE: The Safety Data Sheet (SDS) is available online on the Sakura Finetek USA website at www.sakura-americas.com/SDS.html

CONTACT:

If located within the United States, contact Sakura Finetek USA, Inc. by calling toll free **1-800-725-8723** or contact your Sakura Finetek representative or authorized distributor.

In countries, other than the United States, contact the nearest authorized Sakura Finetek instrument distributor or representative. Contact details may be found at www.sakura.com

REFERENCES:

1. Barrio MM, et al. *Hybridoma* 1998; 17:355-364.
2. Huang CI, et al. *Am J Pathol.* 1998; 153:973-983.
3. Mete O, et al. *Virchows Arch.* 2005; 447:938-946.
4. Azorsa DO, et al. *Blood* 1991; 78:280-284.
5. Woegerbauer M, et al. *Mod Pathol.* 2010; 23:751-762.

SYMBOLS:

- Catalog number
- Batch code
- in vitro* diagnostic medical device
- Temperature limitation
- Use by
- Manufacturer
- Consult instructions for use
- European Conformity
- Authorized representative in the European Community

Storage: 2°C 8°C
Shipping: 2°C 8°C



	Sakura Finetek USA, Inc. 1750 W 214 th Street Torrance, CA 90501 U.S.A.
	Sakura Finetek Europe B.V. Flemingweg 10a 2408 AV Alphen aan den Rijn The Netherlands
Made in U.S.A.	