

Tissue-Tek Genie®

anti-ERG Rabbit Monoclonal Antibody [EP111]

Instructions for use

Intended use

For *in vitro* diagnostic use.

Tissue-Tek Genie® anti-ERG Rabbit Monoclonal Antibody [EP111] is an antibody designed to qualitatively detect ERG protein in formalin-fixed, paraffin embedded (FFPE) specimen sections by immunohistochemistry (IHC) staining on the Tissue-Tek Genie® Advanced Staining System. The clinical interpretation must be made in conjunction with histological examination, relevant clinical information, other diagnostic tests and proper controls by a qualified pathologist.

Limitations

This product has been optimized for use with the default protocol for this antibody on the Tissue-Tek Genie Advanced Staining System, using Tissue-Tek Genie® reagents and FFPE specimen sections. Staining quality may diminish when used with other systems and/or reagents.

Summary and principle

Erythroblast transformation-specific-related (ERG) protein is a transcription factor. ERG protein is involved in hematopoietic and endothelial development. It is expressed in endothelial cells of blood and lymphatic vessels, and bone marrow stem cells. In human prostate cancer, ERG protein is frequently over-expressed due to chromosomal translocations involving ERG and other genes. In particular, the TMPRSS2-ERG fusion gene is found in 45-65% of prostate cancers. TMPRSS2-ERG rearrangement

detected by fluorescent *in situ* hybridization (FISH) have shown a strong correlation with ERG expression detected by IHC. ERG expression has been shown to be a highly specific marker for prostate cancer. ERG is expressed in endothelial neoplasms including hemangioendothelioma, angiosarcoma and Kaposi sarcoma. The anti-ERG antibody is useful for aid in diagnosis of prostate cancer and vascular tumors when used with a panel of other antibodies. Clone EP111 is also named as EPR3864. This clone is capable to detect both wild-type and truncated ERG.

The Tissue-Tek Genie anti-ERG Rabbit Monoclonal Antibody [EP111] is a primary antibody against the ERG protein and is provided in buffered saline containing 1% bovine serum albumin and 0.09% sodium azide. FFPE specimen sections are placed on positively charged slides and the paraffin is removed using the Tissue-Tek Genie® Dewax Solution (REF 8865-G001), after which heat-induced epitope retrieval is performed using the Tissue-Tek Genie® High pH Antigen Retrieval Solution (REF 8744-G001).

IHC demonstration of ERG protein in FFPE specimen sections is achieved through use of the Tissue-Tek Genie anti-ERG Rabbit Monoclonal Antibody [EP111] and the Tissue-Tek Genie® Pro Detection Kit, DAB (REF 8826-K250). This procedure entails the sequential application of antibody and kit components as follows:

- Tissue-Tek Genie® Protein Block
- Tissue-Tek Genie® anti-ERG Rabbit Monoclonal Antibody [EP111]
- Tissue-Tek Genie® Peroxidase Block

- Tissue-Tek Genie® Link (binds to the primary antibody)
- Tissue-Tek Genie® Polymer HRP-Conjugate (binds to the link)
- Tissue-Tek Genie® DAB Substrate (visualizes the detected protein)

Tissue-Tek Genie® Hematoxylin ([REF](#) 8830-M250) is then used to visualize the nuclei of cells. The IHC stained slide is cover-slipped and the FFPE specimen section reviewed using a light microscope.

Expected results

Specificity and intended use of this antibody were validated by performing IHC staining on the Tissue-Tek Genie Advanced Staining System using FFPE normal and tumor specimen sections.

Nuclear staining is observed in endothelial cells of all tissues tested, peripheral T-cells, and mantle zone B-cells of tonsil and appendix. Staining was not observed in epithelial cells and muscle cells in appendix, tonsil, and prostate hyperplasia. Nuclear staining is observed in subset prostate carcinomas.

Sensitivity and identification of ERG protein by this antibody may be affected by improper specimen handling. This may alter antigenicity, weaken detection and may generate false negative results.

Cellular staining pattern: nuclear

Positive specimen control: appendix, tonsil, ERG-positive prostate carcinoma

Cautions and warnings

For professional use only. Take reasonable precautions when handling. Avoid contact of reagents with eyes, skin, and mucous membranes. Wear protective gloves, clothing, and eye/face protection.

Capsules filled with ready-to-use, pre-diluted, antibody are for single use only. Do not attempt to refill or add additional reagent. Discard capsule after use.

Cartridges filled with ready-to-use, pre-diluted, antibody are intended for multiple uses. Do not attempt to refill or add additional reagent. Discard cartridge when empty.

It is recommended to include appropriate controls on each specimen slide to help in identifying any deviation that might occur during the staining process.

All disposal practices must be in compliance with all Federal, State/Provincial and local laws and regulations. Refer to the SDS for further information.

Storage conditions

Store this product at 2-8°C.

Instructions for use

Tissue-Tek Genie® anti-ERG Rabbit Monoclonal Antibody [EP111], capsules ([REF](#) 8323-C010):

1. Place the Tissue-Tek Genie® Reagent Dispensing Area Tag (RDA-Tag) attached to the capsule into the RDA.
2. Push the capsule into the RDA with foil side down and click the attached RDA-Tag down into place on the RDA.
3. Place the RDA on the desired station of the Tissue-Tek Genie Advanced Staining System.
4. Place the slide with the specimen section on the same station, specimen section side down.
5. Assign protocol 8323 to the same station.
6. Initiate execution of protocol 8323.
7. The RDA-Tag 8323 will be scanned and registered automatically when the staining process is initiated.
8. During the primary antibody application step, the antibody will be released from the capsule into the RDA and onto the specimen section on the slide.
9. The staining protocol continues to the end.

Tissue-Tek Genie® anti-ERG Rabbit Monoclonal Antibody [EP111], cartridge ([REF](#) 8323-M250):

1. Prior to placing the cartridge on the carousel of the Tissue-Tek Genie Advanced Staining System, prime the cartridge by facing the nozzle downwards and gently pinching the nozzle tubing until the tubing is filled with the reagent.
2. Place the cartridge on the carousel.
3. Click the RDA-Tag 8323 into place on the RDA.

4. Place the RDA on the desired station of the Tissue-Tek Genie Advanced Staining System.
5. Place the slide with the specimen section on the same station, specimen section side down.
6. Assign protocol 8323 to the same station.
7. Initiate execution of protocol 8323.
8. The RDA-Tag 8323 and the cartridge will be scanned and registered automatically when the staining process is initiated.
9. During the primary antibody application step, the antibody will be dispensed from the cartridge into the RDA and onto the specimen section on the slide.
10. The staining protocol continues to the end.

Material required but not supplied

The following reagents may be required for staining but are not provided:

- Tissue-Tek Genie® Dewax Solution ([REF 8865-G001](#))
- Tissue-Tek Genie® Wash Solution ([REF 8874-G004](#))
- Tissue-Tek Genie® High pH Antigen Retrieval Solution ([REF 8744-G001](#))
- Tissue-Tek Genie® Non-Immune Rabbit Ig Antibody, Negative Control ([REF 8605-C010](#), 8605-M250)
- Tissue-Tek Genie® Pro Detection Kit, DAB ([REF 8826-K250](#))
- Tissue-Tek Genie® Hematoxylin ([REF 8830-M250](#))

Further information can be found on the Sakura Finetek USA website at www.sakuraus.com/Genie

Order information

Product code, product name and quantity

[REF 8323-C010](#) Tissue-Tek Genie® anti-ERG Rabbit Monoclonal Antibody [EP111], Ready-To-Use, 10 capsules; 1 pack.

[REF 8323-M250](#) Tissue-Tek Genie® anti-ERG Rabbit Monoclonal Antibody [EP111], Ready-To-Use, 250 tests, 1 cartridge; 1 unit.

NOTE: The Safety Data Sheet (SDS) is available online on the Sakura Finetek USA website at www.sakuraus.com/SDS.html

References

1. Tomlins SA, et al. Science 2005; 28;310:644-8
2. Petrovics G, et al. Oncogene 2005; 24:3847-3852
3. Perner S, et al. Am J Surg Pathol. 2007 Jun;31(6): 882-8
4. Demichelis F, et al. J Clin Pathol 2007;60:1185-1186
5. Kumar-Sinha et al. Nat Rev Cancer. 2008;8:497-511
6. Park K, et al. Neoplasia 2010;12:590-598
7. van Leenders J, et al. Modern Pathology (2011) 24;1128-1138
8. Tomlins S, et al. Arch Pathol Lab Med. 2012;136: 935-946
9. M Braun, et al. Prostate Cancer and Prostatic Diseases (2012) 15;165-169
10. Yaskiv O, et al. Am J Clin Pathol 2012;138:803-810
11. Tomlins S, et al. Arch Pathol Lab Med. 2012; 136:935-946;
12. Liu H, et al. Annals of Clinical & Laboratory Science 2013;43,3-10
13. Machado I, et al. Pathol Res Pract. 2014; 210:508-13
14. Stockman DL, et al. Mod Pathol. 2014;27:496-501.

Contact

If located within the United States, contact Sakura Finetek USA, Inc. by calling toll free **1-800-725-8723** or contact your Sakura Finetek representative or authorized distributor.

In countries, other than the United States, contact the nearest authorized Sakura Finetek instrument distributor or representative. Contact details may be found at www.sakura.com

Symbols

REF	Catalog number
LOT	Batch code
IVD	<i>in vitro</i> diagnostic medical device
	Temperature limitation
	Use by
	Manufacturer
	Consult instructions for use
	European Conformity
 EC REP	Authorized representative in the European Community

Storage : 2°C



	Sakura Finetek USA, Inc. 1750 W 214 th Street Torrance, CA 90501 U.S.A.
 EC REP	Sakura Finetek Europe B.V. Flemingweg 10a 2408 AV Alphen aan den Rijn The Netherlands
Made in U.S.A.	

GS-33095 Rev. A



continuous innovation for pathology