

# Tissue-Tek Genie®

## anti-Cytokeratin HMW Mouse Monoclonal Antibody [34bE12]

### Instructions for use

#### Intended use

For *in vitro* diagnostic use.

Tissue-Tek Genie® anti-Cytokeratin HMW Mouse Monoclonal Antibody [34bE12] is an antibody designed to qualitatively detect cytokeratin 1, 5, 10, and 14 proteins in formalin-fixed, paraffin embedded (FFPE) specimen sections by immunohistochemistry (IHC) staining on the Tissue-Tek Genie® Advanced Staining System. The clinical interpretation must be made in conjunction with histological examination, relevant clinical information, other diagnostic tests and proper controls by a qualified pathologist.

#### Limitations

This product has been optimized for use with the default protocol for this antibody on the Tissue-Tek Genie Advanced Staining System, using Tissue-Tek Genie® reagents and FFPE specimen sections. Staining quality may diminish when used with other systems and/or reagents.

There is evidence that this antibody cross reacts with an unknown protein, possibly cytokeratin 19 or other cytokeratin. Because of this, it is suggested that this antibody not be used as a general cytokeratin high molecular weight marker. However, evidence of such cross-reaction with this antibody was not seen in prostate epithelial cells.

#### Summary and principle

Human high molecular weight cytokeratins (CK HMW) is a term that collectively refers to the basic cytokeratins 1 through cytokeratin 6 and the acidic

cytokeratins cytokeratin 9 through cytokeratin 17.

Among those, CK HMW clone 34bE12 recognizes cytokeratin 1, 5, 10, and 14. These cytokeratins are expressed in various epithelial basal cells and mesothelial cells. CK HMW clone 34bE12 is useful for differentiating benign prostate gland from prostate adenocarcinoma when used with a panel of antibodies.

Tissue-Tek Genie anti-Cytokeratin HMW Mouse Monoclonal Antibody [34bE12] is a primary antibody against the cytokeratin 1, 5, 10, and 14 proteins and is provided in buffered saline containing 1% bovine serum albumin and 0.09% sodium azide. FFPE specimen sections are placed on positively charged slides and the paraffin is removed using the Tissue-Tek Genie® Dewax Solution (REF 8865-G001), after which heat-induced epitope retrieval is performed using the Tissue-Tek Genie® High pH Antigen Retrieval Solution (REF 8744-G001).

IHC demonstration of cytokeratin 1, 5, 10, and 14 proteins in FFPE specimen sections is achieved through use of the Tissue-Tek Genie anti-Cytokeratin HMW Mouse Monoclonal Antibody [34bE12] and the Tissue-Tek Genie® Pro Detection Kit, DAB (REF 8826-K250). This procedure entails the sequential application of antibody and kit components as follows:

- Tissue-Tek Genie® Protein Block
- Tissue-Tek Genie® anti-Cytokeratin HMW Mouse Monoclonal Antibody [34bE12]
- Tissue-Tek Genie® Peroxidase Block
- Tissue-Tek Genie® Link (binds to the primary antibody)

- Tissue-Tek Genie® Polymer HRP-Conjugate (binds to the link)
- Tissue-Tek Genie® DAB Substrate (visualizes the detected protein)

Tissue-Tek Genie® Hematoxylin ([REF](#) 8830-M250) is then used to visualize the nuclei of cells. The IHC stained slide is cover-slipped and the FFPE specimen section reviewed using a light microscope.

## Expected results

Specificity and intended use of this antibody were validated by performing IHC staining on the Tissue-Tek Genie Advanced Staining System using FFPE normal and tumor specimen sections.

Cytoplasmic staining is observed with stratified squamous epithelia and with basal cells of complex epithelia. The former may be seen in tonsil and esophagus, where cytoplasmic staining is observed in all cell layers of the squamous epithelium. An example of the latter is the cytoplasmic staining of basal cells that is observed in prostatic epithelium (in contrast, secretory cells of the prostate are negative or may show focal staining). Cytoplasmic staining is observed in squamous cell carcinomas of various tissues. Cytoplasmic staining may be observed with breast epithelial cells and neoplastic cells of breast adenocarcinoma; however, this is thought to be a result of the cross-reactivity referenced in the Limitations section.

Sensitivity and identification of cytokeratin 1, 5, 10, and 14 proteins by this antibody may be affected by improper specimen handling. This may alter antigenicity, weaken detection and may generate false negative results.

Cellular staining pattern: cytoplasmic

Positive specimen control: tonsil, esophagus, prostate

## Cautions and warnings

For professional use only. Take reasonable precautions when handling. Avoid contact of reagents with eyes, skin, and mucous membranes. Wear protective gloves, clothing, and eye/face protection.

Capsules filled with ready-to-use, pre-diluted, antibody are for single use only. Do not attempt to refill or add additional reagent. Discard capsule after use.

Cartridges filled with ready-to-use, pre-diluted, antibody are intended for multiple uses. Do not attempt to refill or add additional reagent. Discard cartridge when empty.

It is recommended to include appropriate controls on each specimen slide to help in identifying any deviation that might occur during the staining process.

All disposal practices must be in compliance with all Federal, State/Provincial and local laws and regulations. Refer to the SDS for further information.

## Storage conditions

Store this product at 2-8°C.

## Instructions for use

Tissue-Tek Genie® anti-Cytokeratin HMW Mouse Monoclonal Antibody [34bE12], capsules ([REF](#) 8306-C010):

1. Place the Tissue-Tek Genie® Reagent Dispensing Area Tag (RDA-Tag) attached to the capsule into the RDA.
2. Push the capsule into the RDA with foil side down and click the attached RDA-Tag down into place on the RDA.
3. Place the RDA on the desired station of the Tissue-Tek Genie Advanced Staining System.
4. Place the slide with the specimen section on the same station, specimen section side down.
5. Assign protocol 8306 to the same station.
6. Initiate execution of protocol 8306.
7. The RDA-Tag 8306 will be scanned and registered automatically when the staining process is initiated.
8. During the primary antibody application step, the antibody will be released from the capsule into the RDA and onto the specimen section on the slide.
9. The staining protocol continues to the end.

Tissue-Tek Genie® anti-Cytokeratin HMW Mouse Monoclonal Antibody [34bE12], cartridge ([REF](#) 8306-M250):



- Prior to placing the cartridge on the carousel of the Tissue-Tek Genie Advanced Staining System, prime the cartridge by facing the nozzle downwards and gently pinching the nozzle tubing until the tubing is filled with the reagent.
- Place the cartridge on the carousel.
- Click the RDA-Tag 8306 into place on the RDA.
- Place the RDA on the desired station of the Tissue-Tek Genie Advanced Staining System.
- Place the slide with the specimen section on the same station, specimen section side down.
- Assign protocol 8306 to the same station.
- Initiate execution of protocol 8306.
- The RDA-Tag 8306 and the cartridge will be scanned and registered automatically when the staining process is initiated.
- During the primary antibody application step, the antibody will be dispensed from the cartridge into the RDA and onto the specimen section on the slide.
- The staining protocol continues to the end.

#### **Material required but not supplied**

The following reagents may be required for staining but are not provided:

- Tissue-Tek Genie® Dewax Solution ([REF 8865-G001](#))
- Tissue-Tek Genie® Wash Solution ([REF 8874-G004](#))
- Tissue-Tek Genie® High pH Antigen Retrieval Solution ([REF 8744-G001](#))
- Tissue-Tek Genie® Non-Immune Mouse Ig Antibody, Negative Control ([REF 8604-C010](#), [8604-M250](#))
- Tissue-Tek Genie® Pro Detection Kit, DAB ([REF 8826-K250](#))
- Tissue-Tek Genie® Hematoxylin ([REF 8830-M250](#))

Further information can be found on the Sakura Finetek USA website at [www.sakuraus.com/Genie](http://www.sakuraus.com/Genie)

## **Order information**

### **Product code, product name and quantity**

[REF 8306-C010](#) Tissue-Tek Genie® anti-Cytokeratin HMW Mouse Monoclonal Antibody [34bE12], Ready-To-Use, 10 capsules; 1 pack.

[REF 8306-M250](#) Tissue-Tek Genie® anti-Cytokeratin HMW Mouse Monoclonal Antibody [34bE12], Ready-To-Use, 250 tests, 1 cartridge; 1 unit.

**NOTE:** The Safety Data Sheet (SDS) is available online on the Sakura Finetek USA website at [www.sakuraus.com/SDS.html](http://www.sakuraus.com/SDS.html)

## **References**

- Gown AM, and Vogel AM. J. Cell. Biol. 1982; 95:414-424.
- Gown AM, and Vogel AM. Am. J. Pathol. 1984; 114:309-321.
- O'Malley FP, et al. Virchows Archiv. 1990; 417:191-196.
- Bratthauer GL, et al. J. Histochem. Cytochem. 2003; 51:1527-1531.

## **Contact**

If located within the United States, contact Sakura Finetek USA, Inc. by calling toll free **1-800-725-8723** or contact your Sakura Finetek representative or authorized distributor.

In countries, other than the United States, contact the nearest authorized Sakura Finetek instrument distributor or representative. Contact details may be found at [www.sakura.com](http://www.sakura.com)

## Symbols

<b>REF</b>	Catalog number
<b>LOT</b>	Batch code
<b>IVD</b>	<i>in vitro</i> diagnostic medical device
	Temperature limitation
	Use by
	Manufacturer
	Consult instructions for use
	European Conformity
	Authorized representative in the European Community

Storage : 2°C 8°C



	Sakura Finetek USA, Inc. 1750 W 214 <sup>th</sup> Street Torrance, CA 90501 U.S.A.
	Sakura Finetek Europe B.V. Flemingweg 10a 2408 AV Alphen aan den Rijn The Netherlands
Made in U.S.A.	

GS32079 Rev. A



continuous innovation for pathology