

# Tissue-Tek Genie®

## anti-IgA Rabbit Monoclonal Antibody [EP170]

### Instructions for use

#### Intended use

For *in vitro* diagnostic use.

Tissue-Tek Genie® anti-IgA Rabbit Monoclonal Antibody [EP170] is designed to qualitatively detect IgA proteins in formalin-fixed, paraffin embedded (FFPE) specimen sections by immunohisto-chemistry (IHC) staining on the Tissue-Tek Genie® Advanced Staining System. The clinical interpretation must be made in conjunction with histological examination, relevant clinical information, other diagnostic tests and proper controls by a qualified pathologist.

#### Limitations

This product has been optimized for use with the default protocol for this antibody on the Tissue-Tek Genie Advanced Staining System, using Tissue-Tek Genie® reagents and FFPE specimen sections. Staining quality may diminish when used with other systems and/or reagents.

#### Summary and principle

Immunoglobulin monomers (Igs) are composed of two identical heavy chains and two identical light chains. Variations in the Fc region of the heavy chains distinguishes five different classes of Igs (IgA, IgD, IgE, IgG, and IgM). Each of these classes can be expressed on the B-cell membrane, and more than one heavy chain class can be expressed by the same cell. In normal, healthy serum IgA is generally the second most abundant of the Ig classes. Igs are produced in mature B-cells and plasma cells, and the amount and type of Ig can vary in different stages of B-cell maturation: Igs in

the earlier stages of maturation are present in the cytoplasm, whereas surface Igs are more characteristic of mature B-cells in the mantle zone. Plasma cells lose membranous expression of Igs, but accumulate them at a high level in the cytoplasm and can secrete them into the extracellular space. B-cell derived neoplasms, such as B-cell lymphoma, and plasma cell neoplasms often produce only one type of clonal Ig light chain.

Anti-IgA antibody is useful for classifying B-cell lymphomas and plasma cell neoplasms when used with a panel of other antibodies

The Tissue-Tek Genie anti-IgA Rabbit Monoclonal Antibody [EP170] is a primary antibody against the human IgA protein and is provided in buffered saline containing 1% bovine serum albumin and 0.09% sodium azide. FFPE specimen sections are placed on positively charged slides and the paraffin is removed using the Tissue-Tek Genie® Dewax Solution (REF 8865), after which heat-induced epitope retrieval is performed using the Tissue-Tek Genie® High pH Retrieval Solution (REF 8744).

IHC demonstration of IgA proteins in FFPE specimen sections is achieved through use of the Tissue-Tek Genie anti-IgA Rabbit Monoclonal Antibody [EP170] and the Tissue-Tek Genie® Pro Detection Kit, DAB (REF 8826). This procedure entails the sequential application of antibody and kit components as follows:

- Tissue-Tek Genie® Protein Block
- Tissue-Tek Genie® IgA Rabbit Monoclonal Antibody [EP170]
- Tissue-Tek Genie® Peroxidase Block

- Tissue-Tek Genie® Link (binds to the primary antibody)
- Tissue-Tek Genie® Polymer HRP-Conjugate (binds to the link)
- Tissue-Tek Genie® DAB Substrate (visualizes the detected protein)

Tissue-Tek Genie® Hematoxylin ([REF](#) 8830) is then used to visualize the nuclei of cells. The IHC stained slide is cover-slipped and the FFPE specimen section reviewed using a light microscope.

## Expected results

Specificity and intended use of this antibody were validated by performing IHC staining on the Tissue-Tek Genie Advanced Staining System using FFPE normal and tumor specimen sections. In tonsil and lymph node, staining is observed in the cytoplasm of plasma cells, and in the cytoplasm and plasma membrane of immunoblasts in the germinal center. Some background staining in blood vessels, connective tissue, and epithelial cells may be present. Membranous and cytoplasmic staining is also found in a subset of B-cell neoplasms.

Sensitivity and identification of IgA proteins by this antibody may be affected by improper specimen handling. This may alter antigenicity, weaken detection and may generate false negative results.

Cellular staining pattern: cytoplasmic and membranous staining

Positive control tissue: tonsil, lymph node, selective B-cell neoplasms

## Cautions and warnings

For professional use only. Take reasonable precautions when handling. Avoid contact of reagents with eyes, skin, and mucous membranes. Wear protective gloves, clothing, and eye/face protection.

Capsules filled with ready-to-use, pre-diluted, antibody are for single use only. Do not attempt to refill or add additional reagent. Discard capsule after use.

Cartridges filled with ready-to-use, pre-diluted, antibody are intended for multiple uses. Do not attempt to refill or add additional reagent. Discard cartridge when empty.

It is recommended to include appropriate controls on each specimen slide to help in identifying any deviation that might occur during the staining process.

All disposal practices must be in compliance with all Federal, State/Provincial and local laws and regulations. Refer to the SDS for further information.

## Storage conditions

Store this product at 2-8°C.

## Instructions for use

Tissue-Tek Genie® anti-IgA Rabbit Monoclonal Antibody [EP170], capsules ([REF](#) 8381-C010):

1. Place the Tissue-Tek Genie® Reagent Dispensing Area Tag (RDA-Tag) attached to the capsule into the RDA.
2. Push the capsule into the RDA with foil side down and click the attached RDA-Tag down into place on the RDA.
3. Place the RDA on the desired station of the Tissue-Tek Genie Advanced Staining System.
4. Place the slide with the specimen section on the same station, specimen section side down.
5. Assign protocol 8381 to the same station.
6. Initiate execution of protocol 8381.
7. The RDA-Tag 8381 will be scanned and registered automatically when the staining process is initiated.
8. During the primary antibody application step, the antibody will be released from the capsule into the RDA and onto the specimen section on the slide.
9. The staining protocol continues to the end.

Tissue-Tek Genie® anti-IgA Rabbit Monoclonal Antibody [EP170], cartridge ([REF](#) 8381-M250):

1. Prior to placing the cartridge on the carousel of the Tissue-Tek Genie Advanced Staining System, prime the cartridge by facing the nozzle downwards and gently pinching the nozzle tubing until the tubing is filled with the reagent.
2. Place the cartridge on the carousel.
3. Click the RDA-Tag 8381 into place on the RDA.
4. Place the RDA on the desired station of the Tissue-Tek Genie Advanced Staining System.



5. Place the slide with the specimen section on the same station, specimen section side down.
6. Assign protocol 8381 to the same station.
7. Initiate execution of protocol 8381.
8. The RDA-Tag 8381 and the cartridge will be scanned and registered automatically when the staining process is initiated.
9. During the primary antibody application step, the antibody will be dispensed from the cartridge into the RDA and onto the specimen section on the slide.
10. The staining protocol continues to the end.

#### **Material required but not supplied**

The following reagents may be required for staining but are not provided:

- Tissue-Tek Genie® Dewax Solution ([REF](#) 8865)
- Tissue-Tek Genie® Wash Solution ([REF](#) 8874)
- Tissue-Tek Genie® High pH Antigen Retrieval Solution ([REF](#) 8744)
- Tissue-Tek Genie® Non-Immune Rabbit Ig Antibody, Negative Control ([REF](#) 8605)
- Tissue-Tek Genie® Pro Detection Kit, DAB ([REF](#) 8826)
- Tissue-Tek Genie® Hematoxylin ([REF](#) 8830)

Further information can be found on the  
Sakura Finetek USA website at  
[www.sakuraus.com/Genie](http://www.sakuraus.com/Genie)

#### **Order information**

##### **Product code, product name and quantity**

[REF](#) 8381-C010 Tissue-Tek Genie® anti-IgA Rabbit Monoclonal Antibody [EP170], Ready-To-Use, 10 capsules; 1 pack.

[REF](#) 8381-M250 Tissue-Tek Genie® anti-IgA Rabbit Monoclonal Antibody [EP170], Ready-To-Use, 250 tests, 1 cartridge; 1 unit.

**NOTE:** The Safety Data Sheet (SDS) is available online on the Sakura Finetek USA website at [www.sakuraus.com/SDS](http://www.sakuraus.com/SDS)

#### **References**

1. Leong AS, et al. Manual of Diagnostic Antibodies for Immunohistology. Greenwich Medical Media, Ltd. 1999. London. pp 217-219.
2. Higgins RA, et al. Arch Pathol Lab Med. 2008; 132:441–461.

#### **Contact**

If located within the United States, contact Sakura Finetek USA, Inc. by calling toll free **1-800-725-8723** or contact your Sakura Finetek representative or authorized distributor.

In countries, other than the United States, contact the nearest authorized Sakura Finetek instrument distributor or representative. Contact details may be found at [www.sakura.com](http://www.sakura.com)

## Symbols

<b>REF</b>	Catalog number
<b>LOT</b>	Batch code
<b>IVD</b>	<i>in vitro</i> diagnostic medical device
	Temperature limitation
	Use by
	Manufacturer
	Consult instructions for use
	European Conformity
	Authorized representative in the European Community

Storage: 2°C 8°C



	Sakura Finetek USA, Inc. 1750 W 214 <sup>th</sup> Street Torrance, CA 90501 U.S.A.
	Sakura Finetek Europe B.V. Flemingweg 10a 2408 AV Alphen aan den Rijn The Netherlands
Made in U.S.A.	

GS-33030 Rev.A



continuous innovation for pathology